

SOLUTION CHANGE

ORANGETOWN TOWN
POLICE DEPT.
DATAMASTER dmt: 115206

Date: 04/03/2021

OPERATOR NAME:
JOSEPH W SULLIVAN
Permit #: 46948
Operator Agency/Dept: ORANGETOWN TOWN P.D.
Reference Standard #: 21020

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:00
INTERNAL STANDARD	VERIFIED	21:00
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:01
BLANK TEST	0.00	21:01
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:02
BLANK TEST	0.00	21:03
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:03
BLANK TEST	0.00	21:04

