

SUPERVISOR TEST

ORANGETOWN TOWN
POLICE DEPT.
DATAMASTER dmt: 115206

Date: 04/24/2023

OPERATOR NAME:
JOSEPH W SULLIVAN
Permit #: 46948
Operator Agency/Dept: ORANGETOWN TOWN P.D.
Reference Standard #: 23030

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:42
INTERNAL STANDARD	VERIFIED	15:42
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	15:43
BLANK TEST	0.00	15:43
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	15:44
BLANK TEST	0.00	15:45
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	15:45
BLANK TEST	0.00	15:46

Making a false statement which a person does not believe to be true is a Class A misdemeanor in violation of Section 210.45 of the penal law of the State of New York.

Date 04/24/23

Signature Joseph W Sullivan

