

SUPERVISOR TEST

ORANGETOWN TOWN  
POLICE DEPT.  
DATAMASTER dmt: 115206

Date: 04/25/2022

OPERATOR NAME:  
JOSEPH W SULLIVAN  
Permit #: 46948  
Operator Agency/Dept: ORANGETOWN TOWN P.D.  
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:24
INTERNAL STANDARD	VERIFIED	08:24
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:24
BLANK TEST	0.00	08:25
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	08:26
BLANK TEST	0.00	08:27
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:27
BLANK TEST	0.00	08:28

Making a false statement which a person does not believe to be true is a Class A misdemeanor in violation of Section 210.45 of the penal law of the State of New York.

Date 04/25/2022  
Signature [Handwritten Signature]

