

SUPERVISOR TEST

ORANGETOWN TOWN
POLICE DEPT.
DATAMASTER dmt: 115206

Date: 04/26/2021

OPERATOR NAME:
JOSEPH W SULLIVAN
Permit #: 46948
Operator Agency/Dept: ORANGETOWN TOWN P.D.
Reference Standard #: 21020

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:45
INTERNAL STANDARD	VERIFIED	07:45
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	07:45
BLANK TEST	0.00	07:46
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	07:46
BLANK TEST	0.00	07:47
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	07:48
BLANK TEST	0.00	07:48

