

SUPERVISOR TEST

ORANGETOWN TOWN
POLICE DEPT.
DATAMASTER dmt: 115206

Date: 08/12/2021

OPERATOR NAME:
JOSEPH W SULLIVAN
Permit #: 46948
Operator Agency/Dept: ORANGETOWN TOWN P.D.
Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:55
INTERNAL STANDARD	VERIFIED	07:55
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	07:55
BLANK TEST	0.00	07:56
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	07:57
BLANK TEST	0.00	07:58
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	07:58
BLANK TEST	0.00	07:59

