

SUPERVISOR TEST

ORANGETOWN TOWN  
POLICE DEPT.  
DATAMASTER dmt: 115206

Date: 08/14/2023

OPERATOR NAME:  
JOSEPH W SULLIVAN  
Permit #: 46948  
Operator Agency/Dept: ORANGETOWN TOWN P.D.  
Reference Standard #: 23190

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:12
INTERNAL STANDARD	VERIFIED	08:12
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:13
BLANK TEST	0.00	08:13
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:14
BLANK TEST	0.00	08:15
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:15
BLANK TEST	0.00	08:16

Making a false statement which a person does not believe to be true is a Class A misdemeanor in violation of Section 210.45 of the penal law of the State of New York.

Date 08/14/23  
Signature [Signature]

