

SUPERVISOR TEST

ORANGETOWN TOWN
POLICE DEPT.
DATAMASTER dmt: 115206

Date: 08/18/2024

OPERATOR NAME:
T AMBROSE
Permit #: 53434
Operator Agency/Dept: ORANGETOWN TOWN P.D.
Reference Standard #: 24190

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:58
INTERNAL STANDARD	VERIFIED	04:58
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:59
BLANK TEST	0.00	04:59
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:00
BLANK TEST	0.00	05:01
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:01
BLANK TEST	0.00	05:02

Making a false statement which a person does not believe to be true is a Class A misdemeanor in violation of Section 210.45 of the penal law of the State of New York.

Date 8/18/24
Signature [Signature]

