

SUPERVISOR TEST

ORANGETOWN TOWN
POLICE DEPT.
DATAMASTER dmt: 115206

Date: 08/23/2022

OPERATOR NAME:
JOSEPH W SULLIVAN
Permit #: 46948
Operator Agency/Dept: ORANGETOWN TOWN P.D.
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:05
INTERNAL STANDARD	VERIFIED	08:06
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:06
BLANK TEST	0.00	08:07
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:07
BLANK TEST	0.00	08:08
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:08
BLANK TEST	0.00	08:09

Making a false statement which a person does not believe to be true is a Class A misdemeanor in violation of Section 210.45 of the penal law of the State of New York.

Date 8/23/22
Signature SOB. J. SULLIVAN

