

SUPERVISOR TEST




ORANGETOWN TOWN  
POLICE DEPT.  
DATAMASTER dmt: 115206

Date: 08/30/2022

OPERATOR NAME:  
A PALAZOLO  
Permit #: 42477  
Operator Agency/Dept: ORANGETOWN TOWN P.D.  
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:24
INTERNAL STANDARD	VERIFIED	16:24
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	16:24
BLANK TEST	0.00	16:25
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	16:26
BLANK TEST	0.00	16:26
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	16:27
BLANK TEST	0.00	16:28



Making a false statement which a person does not believe to be true is a Class A misdemeanor in violation of Section 210.45 of the penal law of the State of New York.

08/30/2022

Signature A/567 #228