

SUPERVISOR TEST

ORANGETOWN TOWN
POLICE DEPT.
DATAMASTER dmt: 115206

Date: 12/01/2024

OPERATOR NAME:
T AMBROSE
Permit #: 53435
Operator Agency/Dept: ORANGETOWN TOWN P.D.
Reference Standard #: 24270

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:24
INTERNAL STANDARD	VERIFIED	08:25
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	08:25
BLANK TEST	0.00	08:26
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	08:26
BLANK TEST	0.00	08:27
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	08:27
BLANK TEST	0.00	08:28

Making a false statement which a person does not believe to be true is a Class A misdemeanor in violation of Section 210.45 of the penal law of the State of New York.

Date 12/1/24
Signature *T Ambrose*

