

SUPERVISOR TEST

ORANGETOWN TOWN  
POLICE DEPT.  
DATAMASTER dmt: 115206

Date: 12/04/2023

OPERATOR NAME:  
A PALAZOLO  
Permit #: 42477  
Operator Agency/Dept: ORANGETOWN TOWN P.D.  
Reference Standard #: 23270

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:42
INTERNAL STANDARD	VERIFIED	10:42
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	10:42
BLANK TEST	0.00	10:43
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	10:44
BLANK TEST	0.00	10:45
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	10:45
BLANK TEST	0.00	10:46

Making a false statement which a person  
does not believe to be true is a Class A  
misdemeanor in violation of Section 210.45  
of the penal law of the State of New York.

12/4/23  
~~Signature~~ #22P

