

**SOLUTION CHANGE**

ORANGETOWN TOWN  
 POLICE DEPT.  
 DATAMASTER dmt: 115206

Date: 12/07/2021

OPERATOR NAME:  
 JOSEPH W SULLIVAN  
 Permit #: 46948  
 Operator Agency/Dept: ORANGETOWN TOWN P.D.  
 Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:10
INTERNAL STANDARD	VERIFIED	08:10
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:10
BLANK TEST	0.00	08:11
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:11
BLANK TEST	0.00	08:12
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:13
BLANK TEST	0.00	08:14

