

SUPERVISOR TEST

ORANGETOWN TOWN
POLICE DEPT.
DATAMASTER dmt: 115206

Date: 12/12/2024

OPERATOR NAME:
T AMBROSE
Permit #: 53435
Operator Agency/Dept: ORANGETOWN TOWN P.D.
Reference Standard #: 24270

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:11
INTERNAL STANDARD	VERIFIED	20:11
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	20:11
BLANK TEST	0.00	20:12
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	20:12
BLANK TEST	0.00	20:13
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	20:14
BLANK TEST	0.00	20:15

Making a false statement which a person does not believe to be true is a Class A misdemeanor in violation of Section 210.45 of the penal law of the State of New York.
Date 12/12/24
Signature [Signature] #298

