

SUPERVISOR TEST

ORANGETOWN TOWN  
POLICE DEPT.  
DATAMASTER dmt: 115206

Date: 12/21/2023

OPERATOR NAME:  
A PALAZOLO  
Permit #: 42477  
Operator Agency/Dept: ORANGETOWN TOWN P.D.  
Reference Standard #: 23350

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:49
INTERNAL STANDARD	VERIFIED	15:49
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	15:49
BLANK TEST	0.00	15:50
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	15:50
BLANK TEST	0.00	15:51
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	15:52
BLANK TEST	0.00	15:52

Making a false statement which a person does not believe to be true is a Class A misdemeanor in violation of Section 210.45 of the penal law of the State of New York.

Date 12/21/2023

Signature D/Ser #22P

