

SUPERVISOR TEST

ORANGETOWN TOWN  
POLICE DEPT.  
DATAMASTER dmt: 115206

Date: 12/27/2024

OPERATOR NAME:  
T AMBROSE  
Permit #: 53435  
Operator Agency/Dept: ORANGETOWN TOWN P.D.  
Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:39
INTERNAL STANDARD	VERIFIED	20:39
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	20:39
BLANK TEST	0.00	20:40
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	20:41
BLANK TEST	0.00	20:41
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	20:42
BLANK TEST	0.00	20:43

Making a false statement which a person does not believe to be true is a Class A misdemeanor in violation of Section 210.45 of the penal law of the State of New York.

Date 12/27/24  
Signature T Ambrose #258

