

SUPERVISOR TEST

ORANGETOWN TOWN
POLICE DEPT.
DATAMASTER dmt: 115206

Date: 12/29/2021

OPERATOR NAME:
JOSEPH W SULLIVAN
Permit #: 46948
Operator Agency/Dept: ORANGETOWN TOWN P.D.
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:11
INTERNAL STANDARD	VERIFIED	17:11
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	17:11
BLANK TEST	0.00	17:12
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	17:12
BLANK TEST	0.00	17:13
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	17:14
BLANK TEST	0.00	17:14

