

SUPERVISOR TEST

ORANGETOWN TOWN
POLICE DEPT.
DATAMASTER dmt: 115206

Date: 02/08/2025

OPERATOR NAME:
T AMBROSE
Permit #: 53435
Operator Agency/Dept: ORANGETOWN TOWN P.D.
Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:51
INTERNAL STANDARD	VERIFIED	21:52
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	21:52
BLANK TEST	0.00	21:53
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	21:53
BLANK TEST	0.00	21:54
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	21:54
BLANK TEST	0.00	21:55

Making a false statement which a person does not believe to be true is a Class A misdemeanor in violation of Section 210.45 of the penal law of the State of New York.

Date 2/8/25
Signature [Handwritten Signature] #258

