

SUPERVISOR TEST

ORANGETOWN TOWN  
POLICE DEPT.  
DATAMASTER dmt: 115206

Date: 02/15/2024

OPERATOR NAME:  
A PALAZOLO  
Permit #: 42477  
Operator Agency/Dept: ORANGETOWN TOWN P.D.  
Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:57
INTERNAL STANDARD	VERIFIED	15:57
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	15:58
BLANK TEST	0.00	15:59
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	15:59
BLANK TEST	0.00	16:00
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	16:00
BLANK TEST	0.00	16:01

Making a false statement which a person does not believe to be true is a Class A misdemeanor in violation of Section 210.45 of the penal law of the State of New York.

Date 2/15/2024  
Signature [Signature] # 22A

