

SUPERVISOR TEST

ORANGETOWN TOWN
POLICE DEPT.
DATAMASTER dmt: 115206

Date: 02/25/2023

OPERATOR NAME:
JOSEPH W SULLIVAN
Permit #: 46948
Operator Agency/Dept: ORANGETOWN TOWN P.D.
Reference Standard #: 22420

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:04
INTERNAL STANDARD	VERIFIED	16:04
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	16:04
BLANK TEST	0.00	16:05
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	16:06
BLANK TEST	0.00	16:06
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	16:07
BLANK TEST	0.00	16:08

Making a false statement which a person does not believe to be true is a Class A misdemeanor in violation of Section 210.45 of the penal law of the State of New York.

Date 02/25/23
Signature [Handwritten Signature]

