

SUPERVISOR TEST

ORANGETOWN TOWN
POLICE DEPT.
DATAMASTER dmt: 115206

Date: 02/25/2025

OPERATOR NAME:
T AMBROSE
Permit #: 53435
Operator Agency/Dept: ORANGETOWN TOWN P.D.
Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:14
INTERNAL STANDARD	VERIFIED	16:14
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	16:15
BLANK TEST	0.00	16:16
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	16:16
BLANK TEST	0.00	16:17
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	16:17
BLANK TEST	0.00	16:18

Making a false statement which a person does not believe to be true is a Class A misdemeanor in violation of Section 210.45 of the penal law of the State of New York.

Date 2/25/25
Signature [Signature] 258

