

SUPERVISOR TEST

ORANGETOWN TOWN
 POLICE DEPT.
 DATAMASTER dmt: 115206

Date: 01/04/2025

OPERATOR NAME:
 T AMBROSE
 Permit #: 53435
 Operator Agency/Dept: ORANGETOWN TOWN P.D.
 Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:16
INTERNAL STANDARD	VERIFIED	13:16
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	13:16
BLANK TEST	0.00	13:17
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	13:17
BLANK TEST	0.00	13:18
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	13:19
BLANK TEST	0.00	13:20

Making a false statement which a person does not believe to be true is a Class A misdemeanor in violation of Section 210.45 of the penal law of the State of New York.

Date 1/4/25
 Signature [Signature] 298

