

SUPERVISOR TEST

ORANGETOWN TOWN  
POLICE DEPT.  
DATAMASTER dmt: 115206

Date: 01/16/2023

OPERATOR NAME:  
JOSEPH W SULLIVAN  
Permit #: 46948  
Operator Agency/Dept: ORANGETOWN TOWN P.D.  
Reference Standard #: 22390

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:32
INTERNAL STANDARD	VERIFIED	07:32
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	07:33
BLANK TEST	0.00	07:33
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	07:34
BLANK TEST	0.00	07:35
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	07:35
BLANK TEST	0.00	07:36

Making a false statement which a person does not believe to be true is a Class A misdemeanor in violation of Section 210.45 of the penal law of the State of New York.

Date 01/16/23  
Signature Joseph W Sullivan #194

