

SUPERVISOR TEST

ORANGETOWN TOWN
POLICE DEPT.
DATAMASTER dmt: 115206

Date: 01/23/2023

OPERATOR NAME:
JOSEPH W SULLIVAN
Permit #: 46948
Operator Agency/Dept: ORANGETOWN TOWN P.D.
Reference Standard #: 22390

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:42
INTERNAL STANDARD	VERIFIED	14:42
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	14:42
BLANK TEST	0.00	14:43
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	14:43
BLANK TEST	0.00	14:44
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	14:44
BLANK TEST	0.00	14:45

Making a false statement which a person
does not believe to be true is a Class A
misdemeanor in violation of Section 210.45
of the penal law of the State of New York.

Date: 01/23/2023
Signature: [Signature]

