

**SUPERVISOR TEST**

-----  
ORANGETOWN TOWN  
POLICE DEPT.  
DATAMASTER dmt: 115206  
-----

Date: 01/31/2024

OPERATOR NAME:  
A PALAZOLO  
Permit #: 42477  
Operator Agency/Dept: ORANGETOWN TOWN P.D.  
Reference Standard #: 23350

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:36
INTERNAL STANDARD	VERIFIED	14:36
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	14:37
BLANK TEST	0.00	14:38
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	14:38
BLANK TEST	0.00	14:39
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	14:39
BLANK TEST	0.00	14:40

**Making a false statement which a person does not believe to be true is a Class A misdemeanor in violation of Section 210.45 of the penal law of the State of New York.**

Date 1/31/2024

Signature *[Signature]* #228

