

SUPERVISOR TEST

ORANGETOWN TOWN  
POLICE DEPT.  
DATAMASTER dmt: 115206

Date: 07/12/2022

OPERATOR NAME:  
JOSEPH W SULLIVAN  
Permit #: 46948  
Operator Agency/Dept: ORANGETOWN TOWN P.D.  
Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:02
INTERNAL STANDARD	VERIFIED	16:02
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	16:03
BLANK TEST	0.00	16:04
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	16:04
BLANK TEST	0.00	16:05
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	16:05
BLANK TEST	0.00	16:06

Making a false statement which a person does not believe to be true is a Class A misdemeanor in violation of Section 210.45 of the penal law of the State of New York.

Date 07/12/22  
Signature Joseph W Sullivan

