

SUPERVISOR TEST

ORANGETOWN TOWN  
POLICE DEPT.  
DATAMASTER dmt: 115206

Date: 07/13/2021

OPERATOR NAME:  
JOSEPH W SULLIVAN  
Permit #: 46948  
Operator Agency/Dept: ORANGETOWN TOWN P.D.  
Reference Standard #: 21100

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:03
INTERNAL STANDARD	VERIFIED	08:03
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:03
BLANK TEST	0.00	08:04
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:05
BLANK TEST	0.00	08:05
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:06
BLANK TEST	0.00	08:07

