

**SUPERVISOR TEST**

ORANGETOWN TOWN  
POLICE DEPT.  
DATAMASTER dmt: 115206

Date: 07/23/2024

OPERATOR NAME:  
T AMBROSE  
Permit #: 53435  
Operator Agency/Dept: ORANGETOWN TOWN P.D.  
Reference Standard #: 24140

Nominal = 0.10 g/210L

|                         |          |       |
|-------------------------|----------|-------|
| BLANK TEST              | 0.00     | 10:26 |
| INTERNAL STANDARD       | VERIFIED | 10:26 |
| REFERENCE STANDARD TEMP | 34.02c   |       |
| REFERENCE STANDARD      | 0.09     | 10:26 |
| BLANK TEST              | 0.00     | 10:27 |
| REFERENCE STANDARD TEMP | 34.02c   |       |
| REFERENCE STANDARD      | 0.09     | 10:28 |
| BLANK TEST              | 0.00     | 10:29 |
| REFERENCE STANDARD TEMP | 34.02c   |       |
| REFERENCE STANDARD      | 0.09     | 10:29 |
| BLANK TEST              | 0.00     | 10:30 |

Making a false statement which a person does not believe to be true is a Class A misdemeanor in violation of Section 210.45 of the penal law of the State of New York.

7/23/24  
*[Signature]*

