

SUPERVISOR TEST




ORANGETOWN TOWN
POLICE DEPT.
DATAMASTER dmt: 115206

Date: 07/24/2023

OPERATOR NAME:
JOSEPH W SULLIVAN
Permit #: 46948
Operator Agency/Dept: ORANGETOWN TOWN P.D.
Reference Standard #: 23090

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:42
INTERNAL STANDARD	VERIFIED	16:42
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	16:42
BLANK TEST	0.00	16:43
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	16:44
BLANK TEST	0.00	16:45
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	16:45
BLANK TEST	0.00	16:46



Making a false statement which a person does not believe to be true is a Class A misdemeanor in violation of Section 210.45 of the penal law of the State of New York.

Date: 07/24/23
Signature: [Handwritten Signature]