

SOLUTION CHANGE

ORANGETOWN TOWN
POLICE DEPT.
DATAMASTER dmt: 115206

Date: 07/25/2022

OPERATOR NAME:
JOSEPH W SULLIVAN
Permit #: 46948
Operator Agency/Dept: ORANGETOWN TOWN P.D.
Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:37
INTERNAL STANDARD	VERIFIED	20:37
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	20:37
BLANK TEST	0.00	20:38
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	20:38
BLANK TEST	0.00	20:39
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	20:40
BLANK TEST	0.00	20:41

