

SOLUTION CHANGE

ORANGETOWN TOWN
POLICE DEPT.
DATAMASTER dmt: 115206

Date: 06/05/2021

OPERATOR NAME:
JOSEPH W SULLIVAN
Permit #: 46948
Operator Agency/Dept: ORANGETOWN TOWN P.D.
Reference Standard #: 21100

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:14
INTERNAL STANDARD	VERIFIED	16:14
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	16:15
BLANK TEST	0.00	16:16
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	16:16
BLANK TEST	0.00	16:17
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	16:17
BLANK TEST	0.00	16:18

