

SOLUTION CHANGE




ORANGETOWN TOWN
POLICE DEPT.
DATAMASTER dmt: 115206

Date: 06/06/2024

OPERATOR NAME:
A PALAZOLO
Permit #: 42477
Operator Agency/Dept: ORANGETOWN TOWN P.D.
Reference Standard #: 24140

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 10:26 |
| INTERNAL STANDARD | VERIFIED | 10:26 |
| REFERENCE STANDARD TEMP | 34.02c | |
| REFERENCE STANDARD | 0.09 | 10:27 |
| BLANK TEST | 0.00 | 10:28 |
| REFERENCE STANDARD TEMP | 34.02c | |
| REFERENCE STANDARD | 0.10 | 10:28 |
| BLANK TEST | 0.00 | 10:29 |
| REFERENCE STANDARD TEMP | 34.02c | |
| REFERENCE STANDARD | 0.10 | 10:29 |
| BLANK TEST | 0.00 | 10:30 |



Making a false statement which a person does not believe to be true is a Class A misdemeanor in violation of Section 210.45 of the penal law of the State of New York.

Date 6/6/2024
Signature D/Sen #222