

SUPERVISOR TEST

ORANGETOWN TOWN
POLICE DEPT.
DATAMASTER dmt: 115206

Date: 06/16/2023

OPERATOR NAME:
JOSEPH W SULLIVAN
Permit #: 46948
Operator Agency/Dept: ORANGETOWN TOWN P.D.
Reference Standard #: 23090

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:07
INTERNAL STANDARD	VERIFIED	08:07
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:07
BLANK TEST	0.00	08:08
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:08
BLANK TEST	0.00	08:09
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:09
BLANK TEST	0.00	08:10

Making a false statement which a person does not believe to be true is a Class A misdemeanor in violation of Section 210.45 of the Penal Law of the State of New York.

Date 06/16/23
Signature Joseph W Sullivan

