

SUPERVISOR TEST

ORANGETOWN TOWN  
POLICE DEPT.  
DATAMASTER dmt: 115206

Date: 06/20/2022

OPERATOR NAME:  
JOSEPH W SULLIVAN  
Permit #: 46948  
Operator Agency/Dept: ORANGETOWN TOWN P.D.  
Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:05
INTERNAL STANDARD	VERIFIED	08:05
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:05
BLANK TEST	0.00	08:06
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:06
BLANK TEST	0.00	08:07
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:07
BLANK TEST	0.00	08:08

Making a false statement which a person does not believe to be true is a Class A misdemeanor in violation of Section 210.45 of the penal law of the State of New York.

Date 06/20/2022  
Signature SGT. J. Sullivan #124

