

SUPERVISOR TEST

ORANGETOWN TOWN
POLICE DEPT.
DATAMASTER dmt: 115206

Date: 06/25/2021

OPERATOR NAME:
JOSEPH W SULLIVAN
Permit #: 46948
Operator Agency/Dept: ORANGETOWN TOWN P.D.
Reference Standard #: 21100

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:05
INTERNAL STANDARD	VERIFIED	08:05
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:05
BLANK TEST	0.00	08:06
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:06
BLANK TEST	0.00	08:07
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:07
BLANK TEST	0.00	08:08

