

SUPERVISOR TEST




ORANGETOWN TOWN  
POLICE DEPT.  
DATAMASTER dmt: 115206

Date: 03/07/2022

OPERATOR NAME:  
JOSEPH W SULLIVAN  
Permit #: 46948  
Operator Agency/Dept: ORANGETOWN TOWN P.D.  
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:15
INTERNAL STANDARD	VERIFIED	08:15
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:15
BLANK TEST	0.00	08:16
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	08:16
BLANK TEST	0.00	08:17
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:18
BLANK TEST	0.00	08:19



Making a false statement which a person does not believe to be true is a Class A misdemeanor in violation of Section 210.45 of the penal law of the State of New York.

Date 03/07/22  
Signature [Handwritten Signature]