

SUPERVISOR TEST

ORANGETOWN TOWN
POLICE DEPT.
DATAMASTER dmt: 115206

Date: 03/17/2024

OPERATOR NAME:
A PALAZOLO
Permit #: 42477
Operator Agency/Dept: ORANGETOWN TOWN P.D.
Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:41
INTERNAL STANDARD	VERIFIED	23:41
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	23:41
BLANK TEST	0.00	23:42
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	23:42
BLANK TEST	0.00	23:43
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	23:44
BLANK TEST	0.00	23:45

Making a false statement which a person does not believe to be true is a Class A misdemeanor in violation of Section 210.45 of the penal law of the State of New York.
Date Asst 3/17/2024

Signature *[Signature]*

