

SUPERVISOR TEST

ORANGETOWN TOWN  
POLICE DEPT.  
DATAMASTER dmt: 115206

Date: 03/21/2024

OPERATOR NAME:  
A PALAZOLO  
Permit #: 42477  
Operator Agency/Dept: ORANGETOWN TOWN P.D.  
Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:31
INTERNAL STANDARD	VERIFIED	22:31
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	22:32
BLANK TEST	0.00	22:33
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	22:33
BLANK TEST	0.00	22:34
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:34
BLANK TEST	0.00	22:35

Making a false statement which a person does not believe to be true is a Class A misdemeanor in violation of Section 210.45 of the penal law of the State of New York.

Date 3/21/2024  
Signature [Signature]

