

SUPERVISOR TEST

ORANGETOWN TOWN
POLICE DEPT.
DATAMASTER dmt: 115206

Date: 03/23/2021

OPERATOR NAME:
JOSEPH W SULLIVAN
Permit #: 46948
Operator Agency/Dept: ORANGETOWN TOWN P.D.
Reference Standard #: 20500

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:11
INTERNAL STANDARD	VERIFIED	18:11
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	18:11
BLANK TEST	0.00	18:12
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	18:12
BLANK TEST	0.00	18:13
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	18:13
BLANK TEST	0.00	18:14

