

SUPERVISOR TEST

ORANGETOWN TOWN
POLICE DEPT.
DATAMASTER dmt: 115206

Date: 03/27/2025

OPERATOR NAME:
T AMBROSE
Permit #: 53435
Operator Agency/Dept: ORANGETOWN TOWN P.D.
Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:36
INTERNAL STANDARD	VERIFIED	21:36
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	21:36
BLANK TEST	0.00	21:37
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	21:37
BLANK TEST	0.00	21:38
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:39
BLANK TEST	0.00	21:39

Making a false statement which a person
does not believe to be true is a Class A
misdemeanor in violation of Section 210.45
of the penal law of the State of New York.

Date 5/27/25
Signature [Signature]

