

**SUPERVISOR TEST**

ORANGETOWN TOWN  
POLICE DEPT.  
DATAMASTER dmt: 115206

Date: 05/08/2024

OPERATOR NAME:  
A PALAZOLO  
Permit #: 42477  
Operator Agency/Dept: ORANGETOWN TOWN P.D.  
Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:39
INTERNAL STANDARD	VERIFIED	09:39
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	09:40
BLANK TEST	0.00	09:40
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	09:41
BLANK TEST	0.00	09:42
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	09:42
BLANK TEST	0.00	09:43

Making a false statement which a person does not believe to be true is a Class A misdemeanor in violation of Section 210.45 of the penal law of the State of New York.

Date 5/8/2024

Signature *[Handwritten Signature]*

