

SUPERVISOR TEST

ORANGETOWN TOWN  
POLICE DEPT.  
DATAMASTER dmt: 115206

Date: 05/10/2022

OPERATOR NAME:  
JOSEPH W SULLIVAN  
Permit #: 46948  
Operator Agency/Dept: ORANGETOWN TOWN P.D.  
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:36
INTERNAL STANDARD	VERIFIED	07:36
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	07:37
BLANK TEST	0.00	07:38
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	07:38
BLANK TEST	0.00	07:39
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	07:39
BLANK TEST	0.00	07:40

Making a false statement which a person does not believe to be true is a Class A misdemeanor in violation of Section 210.45 of the penal law of the State of New York.

Date 05/10/22

Signature [Signature]

