

SOLUTION CHANGE

ORANGETOWN TOWN
POLICE DEPT.
DATAMASTER dmt: 115206

Date: 05/12/2021

OPERATOR NAME:
JOSEPH W SULLIVAN
Permit #: 46948
Operator Agency/Dept: ORANGETOWN TOWN P.D.
Reference Standard #: 21020

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:56
INTERNAL STANDARD	VERIFIED	06:56
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	06:56
BLANK TEST	0.00	06:57
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	06:57
BLANK TEST	0.00	06:58
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	06:59
BLANK TEST	0.00	06:59

