

SUPERVISOR TEST

ORANGETOWN TOWN
POLICE DEPT.
DATAMASTER dmt: 115206

Date: 05/18/2021

OPERATOR NAME:
JOSEPH W SULLIVAN
Permit #: 46948
Operator Agency/Dept: ORANGETOWN TOWN P.D.
Reference Standard #: 21020

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:08
INTERNAL STANDARD	VERIFIED	19:08
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	19:09
BLANK TEST	0.00	19:09
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	19:10
BLANK TEST	0.00	19:11
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	19:11
BLANK TEST	0.00	19:12

