

SUPERVISOR TEST

ORANGETOWN TOWN
POLICE DEPT.
DATAMASTER dmt: 115206

Date: 05/23/2023

OPERATOR NAME:
JOSEPH W SULLIVAN
Permit #: 46948
Operator Agency/Dept: ORANGETOWN TOWN P.D.
Reference Standard #: 23030

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:45
INTERNAL STANDARD	VERIFIED	15:45
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	15:46
BLANK TEST	0.00	15:47
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	15:47
BLANK TEST	0.00	15:48
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	15:48
BLANK TEST	0.00	15:49

Making a false statement which a person does not believe to be true is a Class A misdemeanor in violation of Section 210.45 of the penal law of the State of New York.

Date 5/23/23
Signature SGT J. Sullivan

