

SUPERVISOR TEST

ORANGETOWN TOWN
POLICE DEPT.
DATAMASTER dmt: 115206

Date: 05/30/2024

OPERATOR NAME:
A PALAZOLO
Permit #: 42477
Operator Agency/Dept: ORANGETOWN TOWN P.D.
Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:33
INTERNAL STANDARD	VERIFIED	15:33
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	15:34
BLANK TEST	0.00	15:35
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	15:35
BLANK TEST	0.00	15:36
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	15:36
BLANK TEST	0.00	15:37

Making a false statement which a person does not believe to be true is a Class A misdemeanor in violation of Section 210.45 of the penal law of the State of New York.

Date 5/30/2024 #228
signature D/SO

