

SOLUTION CHANGE

ORANGETOWN TOWN
POLICE DEPT.
DATAMASTER dmt: 115206

Date: 11/02/2021

OPERATOR NAME:
JOSEPH W SULLIVAN
Permit #: 46948
Operator Agency/Dept: ORANGETOWN TOWN P.D.
Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:57
INTERNAL STANDARD	VERIFIED	15:58
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	15:58
BLANK TEST	0.00	15:59
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	15:59
BLANK TEST	0.00	16:00
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	16:00
BLANK TEST	0.00	16:01

