

SUPERVISOR TEST

ORANGETOWN TOWN
POLICE DEPT.
DATAMASTER dmt: 115206

Date: 11/12/2024

OPERATOR NAME:
T AMBROSE
Permit #: 53435
Operator Agency/Dept: ORANGETOWN TOWN P.D.
Reference Standard #: 24270

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:41
INTERNAL STANDARD	VERIFIED	16:41
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	16:42
BLANK TEST	0.00	16:43
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	16:43
BLANK TEST	0.00	16:44
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	16:44
BLANK TEST	0.00	16:45

Making a false statement which a person does not believe to be true is a Class A misdemeanor in violation of Section 210.45 of the penal law of the State of New York.

Date 11/12/24
Signature Sgt Ambrose 258

