

SUPERVISOR TEST

ORANGETOWN TOWN  
POLICE DEPT.  
DATAMASTER dmt: 115206

Date: 11/16/2023

OPERATOR NAME:  
A PALAZOLO  
Permit #: 42477  
Operator Agency/Dept: ORANGETOWN TOWN P.D.  
Reference Standard #: 23270

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:18
INTERNAL STANDARD	VERIFIED	18:18
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	18:19
BLANK TEST	0.00	18:19
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	18:20
BLANK TEST	0.00	18:21
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	18:21
BLANK TEST	0.00	18:22

Making a false statement which a person does not believe to be true is a Class A misdemeanor in violation of Section 210.45 of the penal law of the State of New York.

Date 11/16/23

Signature D/Sci #228

