

SUPERVISOR TEST

ORANGETOWN TOWN
POLICE DEPT.
DATAMASTER dmt: 115206

Date: 11/17/2022

OPERATOR NAME:
JOSEPH W SULLIVAN
Permit #: 46948
Operator Agency/Dept: ORANGETOWN TOWN P.D.
Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:57
INTERNAL STANDARD	VERIFIED	07:57
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	07:58
BLANK TEST	0.00	07:59
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	07:59
BLANK TEST	0.00	08:00
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:00
BLANK TEST	0.00	08:01

Making a false statement which a person does not believe to be true is a Class A misdemeanor in violation of Section 210.45 of the penal law of the State of New York.

Date 11/17/2022
Signature [Handwritten Signature]

