

SUPERVISOR TEST

ORANGETOWN TOWN
 POLICE DEPT.
 DATAMASTER dmt: 115206

Date: 11/22/2021

OPERATOR NAME:
 JOSEPH W SULLIVAN
 Permit #: 46948
 Operator Agency/Dept: ORANGETOWN TOWN P.D.
 Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:11
INTERNAL STANDARD	VERIFIED	08:11
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:11
BLANK TEST	0.00	08:12
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:13
BLANK TEST	0.00	08:13
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:14
BLANK TEST	0.00	08:15

